

South Carolina Law Enforcement Division - Status Records



Date Range From 01/01/2009 To 04/15/2010
 DataMaster (BAC/DMT) Serial Number - 110907
 Report Date - 4/15/2010

Legend I - Incomplete
 R - Refusal
 V - Invalid Sample
 X - Interference Detected

<u>Serial Number</u>	<u>Loc Number</u>	<u>Date/Time</u>	<u>Status Record</u>	<u>Status Info</u>
110907	SC0100100	2/19/2009 12:55:38 AM	PUMP ERROR	
110907	SC0100100	2/28/2009 12:06:25 AM	FILTER WHEEL ERROR	
110907	SC0100100	3/6/2009 10:45:15 PM	FILTER WHEEL ERROR	
110907	SC0100100	3/9/2009 11:35:42 PM	FILTER WHEEL ERROR	
110907	SC0100100	4/12/2009 12:34:24 AM	INTERFERENCE DETECTED	
110907	SC0100100	5/8/2009 11:21:09 PM	BLANK ERROR	
110907	SC0100100	5/21/2009 6:19:39 PM	INTERFERENCE DETECTED	
110907	SC0100100	6/13/2009 1:43:52 AM	COMMUNICATION ERROR	
110907	SC0100100	6/24/2009 12:35:22 AM	PUMP ERROR	
110907	SC0100100	8/11/2009 3:02:56 AM	INVALID SAMPLE DETECTED	
110907	SC0100100	10/17/2009 10:10:59 PM	COMMUNICATION ERROR	
110907	SC0100100	1/6/2010 2:23:37 AM	INVALID SAMPLE DETECTED	
110907	SC0100100	1/11/2010 1:56:28 AM	INVALID SAMPLE DETECTED	
110907	SC0100100	3/7/2010 2:58:02 AM	EXTERNAL STANDARD OUT OF RANGE	
110907	SC0100100	3/10/2010 2:35:42 AM	EXTERNAL STANDARD OUT OF RANGE	
110907	SC0100100	3/31/2010 10:15:48 PM	INVALID SAMPLE DETECTED	

THIS REPORT REPRESENTS TRUE AND CORRECT COPIES OF THE ORIGINAL DOCUMENTS MAINTAINED BY SLED IN THE NORMAL COURSE OF BUSINESS.

**DEBORAH H. BANKS, LIEUTENANT
 SUPERVISORY SPECIAL AGENT
 IMPLIED CONSENT DEPARTMENT**